




Cobalt / Cobalto




Cat. No.	Cuvette	Measurement range	Method	Wavelength
918 51	10 mm	0.02 - 0.70 mg/l Co ²⁺	(1) 511	540 nm
	50 mm	0.002 - 0.300 mg/l Co ²⁺		




1 20 ml sample
échantillon
muestra




2 1.0 ml R1




3 mix
mélanger
mezclar




4 1.0 ml R2




5 mix
mélanger
mezclar




6 1.0 ml R3




7 mix
mélanger
mezclar




8 fill up with dist. water
emplir avec l'eau dist.
llenar con agua dest.




9 mix
mélanger
mezclar




10 5'00 min




11 rinse and fill
rincer et emplir
lavar y llenar



12 measurement
mesure
medición



blank value
zéro
cero



blank measurement
zéro mesure
cero medición