




Zinc




Cat. No.	Cuvette	Measurement range	Method	Wavelength
918 95	10 mm	0.1 - 3.0 mg/L Zn ²⁺	(1) 951	620 nm
	50 mm	0.02 - 1.50 mg/L Zn ²⁺		




1
20 mL sample
échantillon
muestra




2
1.0 mL R1




3
mix
mélanger
mezclar




4
1.0 mL R2




5
mix
mélanger
mezclar




6
pH 10.5 - 11.5
pH control
pH contrôle
pH control




7
2'00 min




8
1.0 mL R3




9
do not mix
non melangér
tampoco mezclar




10
fill up with dist. water
emplir avec l'eau dist.
llenar con agua dest.




11
mix
mélanger
mezclar




12
rinse and fill
rincer et emplir
lavar y llenar




13
1'00 min



14
M
measurement
mesure
medición



14
N
blank value
zéro
cero



14
N
blank measurement
zéro mesure
cero medición