

# visocolor® ECO Ammonium 3 Amonio

## Test 5-08

REF 931 208

690 nm

Method(e) / Método

5081


0.1–2.0 mg/L NH<sub>4</sub>-N

5082

0.1–2.5 mg/L NH<sub>4</sub><sup>+</sup>

5083


0.1–2.0 mg/L NH<sub>3</sub>




Probe / sample  
échantillon / muestra



Küvetten spülen  
Rinse test tubes  
Rincer les éprouvettes  
Lavar los tubos




5 mL Probe  
Sample  
Echantillon  
Muestra



10 ∩ NH<sub>4</sub>-1




Schütteln  
Shake  
Agiter  
Agitar




1 ∩ NH<sub>4</sub>-2



Schütteln  
Shake  
Agiter  
Agitar




5'00 min




4 ∩ NH<sub>4</sub>-3



Schütteln  
Shake  
Agiter  
Agitar



Säubern  
Clean  
Nettoyer  
Limpiar



7'00 min



Messung  
Measurement  
Mesure  
Medición



1+9  
Meerwasser  
Sea water  
Eau de mer  
Agua de mar