

# visocolor® ECO Brom (frei)

Bromine (free) / Brome (libre) / Bromo (libre)

**Test 5-11**


REF 931 211

540 nm

Method(e) / Método

5111


0.10–13.00 mg/L Br<sub>2</sub>




Probe / sample  
échantillon / muestra



**1**  
Küvetten spülen  
Rinse test tubes  
Rincer les éprouvettes  
Lavar los tubos




**2** 5 mL Probe  
Sample  
Echantillon  
Muestra




**3** 1 Br<sub>2</sub>-1



**4** 20 s Schütteln  
Shake  
Agiter  
Agitar




**5**  
Säubern  
Clean  
Nettoyer  
Limpiar



**6** 1'00 min



**7**  
Messung  
Measurement  
Mesure  
Medición



Meerwasser  
Sea water  
Eau de mer  
Agua de mar

# visicolor<sup>®</sup> ECO Brom (gesamt)

Bromine (total) / Brome (totale) / Bromo (total)

**Test 5-11**


REF 931 211

540 nm


Method(e) / Método

5112


0.10–13.00 mg/L Br<sub>2</sub>




Probe / sample  
échantillon / muestra



1  
Küvetten spülen  
Rinse test tubes  
Rincer les éprouvettes  
Lavar los tubos




2  
5 mL Probe  
Sample  
Echantillon  
Muestra



3  
1  Br<sub>2</sub>-1




4  
20 s Schütteln  
Shake  
Agiter  
Agitar




5  
3  Br<sub>2</sub>-2



6  
Schütteln  
Shake  
Agiter  
Agitar




7  
Säubern  
Clean  
Nettoyer  
Limpiar



8  
2'00 min



9  
Messung  
Measurement  
Mesure  
Medición



Meerwasser  
Sea water  
Eau de mer  
Agua de mar