

visicolor® ECO Nitrit

Nitrite / Nitrito

Test 5-44

REF 931 244

540 nm


Method(e) / Método

5441


0.01–0.15 mg/L NO₂-N

5442


0.02–0.50 mg/L NO₂⁻




Probe / sample
échantillon / muestra




1
Küvetten spülen
Rinse test tubes
Rincer les éprouvettes
Lavar los tubos




2
5 mL Probe
Sample
Echantillon
Muestra



3
4  NO₂-1




4
Schütteln
Shake
Agiter
Agitar




5
1  NO₂-2



6
Schütteln
Shake
Agiter
Agitar




7
Säubern
Clean
Nettoyer
Limpiar



8
10'00 min



9
Messung
Measurement
Mesure
Medición



Meerwasser
Sea water
Eau de mer
Agua de mar