


NANOCOLOR[®] Chromat 5

Chromate / Cromato



Test 0-24
REF 985024

Methode / Method / Méthode / Método

0241 0.05–2.00 mg/L Cr(VI)

0242 0.1–4.0 mg/L CrO₄²⁻

1241  0.005–0.500 mg/L Cr(VI)

1242  0.01–1.00 mg/L CrO₄²⁻

540 nm

Min	20–80 %		Max
0.05	0.44	1.61	2.00 mg/L Cr(VI)
0.1	0.88	3.22	4.0 mg/L CrO ₄ ²⁻

