


NANOCOLOR® Molybdän 40

Molybdenum 40 / Molybdène 40 / Molibdeno 40



Test 0-56
REF 985056

Methode / Method / Méthode / Método

0561 1,0–40,0 mg/L Mo(VI)

0562 1,6–65,0 mg/L MoO₄²⁻

345 / 365 nm

| Min | 20–80 % | | Max |
|-----|---------|------|-----------|
| 1,0 | 9,0 | 32,0 | 40,0 mg/L |
| 1,6 | 14,3 | 52,3 | 65,0 mg/L |

