


NANOCOLOR[®] gesamt-Phosphat 50

total Phosphate / Phosphate total / Fosfato total













Test 0-79
REF 985079

Methode / Method / Méthode / Método

0791 10.0–50.0 mg/L P
0792 30–150 mg/L PO₄³⁻
0793 23–110 mg/L P₂O₅


436 nm

Min	20–80 %		Max
10.0	18.0	42.0	50.0 mg/L P
30	54	126	150 mg/L PO ₄ ³⁻



NANOCOLOR[®] ortho-Phosphat 50

ortho-Phosphate / ortho Phosphates / orto Fosfato



Test 0-79
REF 985079

Methode / Method / Méthode / Método

0794 10.0–50.0 mg/L PO₄-P

0795 30–150 mg/L PO₄³⁻

0796 23–110 mg/L P₂O₅

436 nm

Min	20–80 %		Max
10.0	18.0	42.0	50.0 mg/L PO ₄ -P
30	54	126	150 mg/L PO ₄ ³⁻

